

EC Declaration of Conformity

Manufacturer: *DeVilbiss Healthcare LLC
100 DeVilbiss Drive
Somerset, PA 15501, USA*

EC Authorized Representative: *DeVilbiss Healthcare GmbH
Kamenzerstraße 3, 68309
Mannheim, Germany*

1. Oxygen Delivery Units, controlled (UMDNS 18-076):

Catalogue nos.: *535I*

Classification (MDD Annex IX): *IIa (Rule 11)*
Conformity Assessment Procedure: *MDD 93/42/EEC, Annex II excluding Section 4*

2. Accessories:

Product Description (Catalogue no.):

<i>ML6 Cylinder w/CF Regulator, Europe</i>	<i>535I-ML6-CF</i>
<i>C Cylinder w/CF Regulator, Europe</i>	<i>535I-C-CF</i>
<i>D Cylinder w/CF Regulator, Europe</i>	<i>535I-D-CF</i>
<i>E Cylinder w/CF Regulator, Europe</i>	<i>535I-E-CF</i>
<i>Cabinet Air Filter, 6 pk.</i>	<i>535D-605</i>
<i>Bag, C Cylinder</i>	<i>EX3000D-651</i>
<i>Bag, D Cylinder</i>	<i>EX3000D-652</i>
<i>Bag, M-6 Mini Cylinder</i>	<i>EX3000D-653</i>
<i>Bag, ML6 Cylinder</i>	<i>EX3000D-654</i>
<i>Cylinder Nipple Dust Cap, 5 pk.</i>	<i>PD1000A-627</i>
<i>Covers, Top & Bottom replacement</i>	<i>PD1000A-601</i>
<i>Gauge Kit</i>	<i>PD1000A-620</i>
<i>Gauge Kit</i>	<i>PD1000G-620</i>
<i>Seal, Regulator, 10 pk.</i>	<i>PD1000G-618</i>
<i>PD1000A w/C Cylinder, Europe</i>	<i>PD1000A-I-C</i>
<i>PD1000A w/D Cylinder, Europe</i>	<i>PD1000A-I-D</i>
<i>PD1000A w/E Cylinder, Europe</i>	<i>PD1000A-I-E</i>
<i>PD1000A w/ML6 Cylinder, Europe</i>	<i>PD1000A-I-ML6</i>

Applied standards: All applicable harmonized standards as adopted under the EC Directive 93/42/EEC and published in the Official Journal of the European Communities. *(See attached listing)*

This Declaration of Conformity is based on the EC Directive 93/42/EEC, Annex II.3 and supported by the TÜV NORD CERT GmbH (0044) certificate, with reference to articles 1 and 3 of the directive 93/42/EEC.

Notified Body: *TÜV NORD CERT GmbH
Langemarckstrasse 20, 45141 Essen, Germany*

Identification No.: *0044*
EC Certificate No.: *44 232 117803*
Start of EC Marking: *2002-06-19*

EC Declaration of Conformity

We, DeVilbiss Healthcare LLC., hereby declare under our sole responsibility that the above-mentioned products comply with the essential requirements and provisions of Council Directive 93/42/EEC.



Somerset, PA, 30 July 2024

Place Date

 Zita A. Yurko, RAC
 Sr. Director Regulatory Affairs
 Name and Position

Applied Standards:

535 series
<i>IEC 60601-1:2005+A1:2012 Medical electrical equipment — Part 1 General requirements for basic safety</i>
<i>IEC 60601-1-2:2014, Medical electrical equipment — Part 1-2: General requirements for basic safety and essential performance — Collateral Standard: Electromagnetic compatibility – Requirements and tests (FDA Recognition Number 19-8)</i>
<i>IEC 60601-1-6:2010 + AMD 1:2013 Ed. 3.1 Medical electrical equipment - Part 1-6: General requirements for basic safety and essential performance - Collateral standard: Usability (associated with IEC 60601-1 Ed. 3.0) (FDA Recognition Number 5-89)</i>
<i>IEC 60601-1-9:2007 + A1:2013 Ed. 1.1 Medical electrical equipment - Part 1-9: General requirements for basic safety and essential performance - Collateral standard: Requirements for environmentally conscious design (associated with IEC 60601-1 Ed. 3.0)</i>
<i>IEC 60601-1-11:2015 (Ed 2.0), Medical electrical equipment — Part 1-11: General requirements for basic safety and essential performance – Collateral Standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment (FDA Recognition Number 19-14)</i>
<i>IEC 62366:2007 Ed. 1.0 + AMD 1:2014 – Medical devices - Application of usability engineering to medical devices (FDA Recognition number 5-87)</i>
<i>IEC 62304:2006 + AMD1:2015: Ed. 1.1 - Medical device software - Software life cycle processes (FDA Recognition number 13-79)</i>
<i>BS EN ISO 14971:2012 Medical devices - Application of risk management to medical devices (FDA Recognition number 5-40 is to ISO 14971:2007)</i>
<i>ISTA Procedure 3A: Packaged-Products for Parcel Delivery System Shipments 70kg (150 lb) or less (standard).</i>
<i>ASTM G63-15 Standard Guide for Evaluating Nonmetallic Materials for Oxygen Service</i>
<i>ASTM G94-05 Standard Guide for Evaluating Metals for Oxygen Service</i>
<i>BS EN ISO 15001:2011 Anaesthetic and respiratory equipment. Compatibility with oxygen</i>